

产品中有害物质的名称及含量

Name and Concentration of Hazardous Substances in Product

Atellica® IM 1300/1600 Analyzer

部件名称 Part Name	有害物质 Hazardous Substance					
	铅 (Pb)	汞 (Hg)	镉 (Cd)	六价铬 (Cr (VI))	多溴联苯 (PBB)	多溴二苯醚 (PBDE)
盖子 Covers	X	0	0	0	0	0
反应杯加载器 Cuvette loader	X	0	0	0	0	0
设备控制模块 Device control module	X	0	0	0	0	0
加热风机 Heater fan	X	0	0	0	0	0
电源 Power supplies	X	0	X	0	0	0
试剂仓湿度传感器 Reagent compartment humidity sensor	X	0	0	0	0	0
样本吸头装载机 Sample tip loader	X	0	0	0	0	0

部件名称 Part Name	有害物质 Hazardous Substance					
	铅 (Pb)	汞 (Hg)	镉 (Cd)	六价铬 (Cr (VI))	多溴联苯 (PBB)	多溴二苯醚 (PBDE)
固体废弃物仓 Solid waste compartment	x	o	o	o	o	o
供水和真空压力组件 Water and vacuum pressure assembly	x	o	o	o	o	o

本表格依据 SJ/T 11364 的规定编制。

This table is prepared in accordance with the provisions of SJ/T 11364.

o	表示该有害物质在该部件所有均质材料中的含量均在 GB/T 26572 规定的限量要求以下。 Indicates that said hazardous substance contained in all of the homogeneous materials for this part is below the limit requirement of GB/T 26572.
x	表示该有害物质至少在该部件的某一均质材料中的含量超出 GB/T 26572 规定的限量要求。 Indicates that said hazardous substance contained in at least one of the homogeneous materials used for this part is above the limit requirement of GB/T 26572.



除非另外特别的标注,此标志为西门子医疗产品及其部件的环保使用期。此环保使用期限只适用于产品在产品手册中所规定的条件下工作。

The Environmentally Friendly Use Period (EFUP) for Products of Siemens Healthcare and their parts is per the symbol shown here, unless otherwise marked. The EFUP is valid only when the product is operated under the conditions defined in the user documentation.

商标信息

Atellica 是 Siemens Healthcare Diagnostics Inc 的商标。