

Syphilis Quality Control (Syph QC)

Current Revision and Date ^a	Rev. 02, 2019-09
Product Name	Atellica IM Syphilis Quality Control (Syph QC)
Abbreviated Product Name	Atellica IM Syph QC
	2 x 7.0 mL negative quality control CONTROL - REF 10995676 2 x 7.0 mL positive quality control CONTROL + Quality control lot-specific value sheet CONTROL LOT VAL
Systems	Atellica IM Analyzer

^a A vertical bar in the page margin indicates technical content that differs from the previous version.



Intended Use

The Atellica™ IM Syphilis Quality Control (Syph QC) is for *in vitro* diagnostic use in monitoring the performance of the Atellica IM Syph assay using the Atellica™ IM Analyzer.

Material Description

Material Description	Storage	Stability ^a
Atellica IM Syph QC 7.0 mL/vial Processed human plasma negative and positive for <i>Treponema pallidum</i> antibodies; preservatives	At 2–8°C Onboard at room temperature Atellica™ Sample Handler ^b	Until expiration date on product 8 hours

^a Refer to *Storage and Stability*.

^b Refer to the supplementary document "Atellica Sample Handler Calibrator and QC Storage and Stability" for information about storage and stability of materials in the Cal-QC tube storage area.

Warnings and Precautions

For *in vitro* diagnostic use.

For Professional Use.

CAUTION

Federal (USA) law restricts this device to sale by or on the order of a licensed healthcare professional.

Safety data sheets (SDS) available on siemens.com/healthineers.

H412 P273, P501	Harmful to aquatic life with long lasting effects. Avoid release to the environment. Dispose of contents and container in accordance with all local, regional, and national regulations. Contains: 2-methyl-2H-isothiazol 3-one
----------------------------------	--



CAUTION POTENTIAL BIOHAZARD

Contains human source material. Each donation of human blood or blood component was tested by FDA-approved methods for the presence of antibodies to human immunodeficiency virus type 1 (HIV-1) and type 2 (HIV-2), as well as for hepatitis B surface antigen (HBsAg) and antibody to hepatitis C virus (HCV). The test results were negative (not repeatedly reactive). The positive control contains processed human plasma that is reactive for antibody to *T. pallidum*. No test offers complete assurance that these or other infectious agents are absent; this material should be handled using good laboratory practices and universal precautions.¹⁻³

Dispose of hazardous or biologically contaminated materials according to the practices of your institution. Discard all materials in a safe and acceptable manner and in compliance with prevailing regulatory requirements.

Note For information about quality control material preparation, refer to *Preparing the Quality Control Materials*.

Storage and Stability

Store quality control materials in an upright position. Quality control materials are stable until the expiration date on the product when stored at 2–8°C. Quality control materials are stable for 8 hours on the system at room temperature.

Refer to the supplementary document "Atellica Sample Handler Calibrator and QC Storage and Stability" for information about storage and stability of materials in the Cal-QC tube storage area.

Do not use products beyond the expiration date printed on the product labeling.

Performing Quality Control

Perform the quality control procedure at least once during each work shift that samples are analyzed.

Follow government regulations or accreditation requirements for quality control frequency. Individual laboratory quality control programs and procedures may require more frequent quality control testing.

Test quality control samples after a successful calibration.

Preparing the Quality Control Materials

Quality control materials are liquid and ready to use. Gently mix and invert the vials to ensure homogeneity of the material.

Note Use quality control material within the stability limits specified in *Storage and Stability* and discard any remaining material.

Quality Control Procedure

The required sample volume for testing depends on several factors. For information about sample volume requirements, refer to the online help.

Use the following lot-specific materials to perform quality control:

- For the quality control (QC) definitions, refer to the lot-specific value sheet

CONTROL	LOT	VAL
---------	-----	-----

 provided with the quality control materials.
- Generate lot-specific barcode labels to use with the quality control samples.

For instructions about how to perform the quality control procedure, refer to the online help.

Taking Corrective Action

If the quality control results do not fall within the assigned values, do not report results. Perform corrective actions in accordance with established laboratory protocol. For suggested protocol, refer to the online help.

Expected Values

For the assigned values, refer to the quality control lot-specific value sheet

CONTROL	LOT	VAL
---------	-----	-----

 provided. A satisfactory level of performance is achieved when the analyte values obtained are within the expected control range for the system or within your range, as determined by an appropriate internal laboratory quality control scheme. Follow your laboratory's quality control procedures if the results obtained do not fall within the acceptable limits. For information about entering QC definitions, refer to the online help.

The assigned values are traceable to the standardization of the assay. For additional information, refer to the assay instructions for use.

Limitations

The performance of the Atellica IM Syph QC has not been established with any other syphilis assay.

The results obtained using quality control material depend on several factors. Erroneous results can occur from causes such as improper storage, inadequate mixing, reconstitution errors, or sample handling errors associated with system or assay procedures.

The assigned control values should be used as a guide in evaluating performance. The control intervals and ranges should be adapted to each laboratory's individual requirements. Values obtained should fall within the established interval. Each laboratory should establish corrective measures if individual values fall outside the interval. Follow the applicable government regulations and local guidelines for quality control.

Technical Assistance

For customer support, contact your local technical support provider or distributor.
siemens.com/healthineers



















References









1. Centers for Disease Control. Perspectives in disease prevention and health promotion update: Universal precautions for prevention of transmission of human immunodeficiency virus, hepatitis B virus and other bloodborne pathogens in healthcare settings. *MMWR*. 1988;37(24):377–382, 387–388.
2. Clinical and Laboratory Standards Institute. *Procedures for the Handling and Processing of Blood Specimens for Common Laboratory Tests; Approved Guideline—Fourth Edition*. Wayne, PA: Clinical and Laboratory Standards Institute; 2010. CLSI Document GP44-A4.
3. Clinical and Laboratory Standards Institute. *Protection of Laboratory Workers From Occupationally Acquired Infections; Approved Guideline—Fourth Edition*. Wayne, PA: Clinical and Laboratory Standards Institute; 2014. CLSI Document M29-A4.

Definition of Symbols

The following symbols may appear on the product labeling:

Symbol	Symbol Title and Description
	Consult instructions for use
 Rev. 01	Version of instructions for use
siemens.com/healthcare	Internet URL address to access the electronic instructions for use
siemens.com/document-library	
Rev.	Revision
	Caution Consult instructions for use or accompanying documents for cautionary information such as warnings and precautions that cannot, for a variety of reasons, be presented on the medical device.
	Biological risks Potential biological risks are associated with the medical device.
	Corrosive
	Dangerous to environment
	Irritant Oral, dermal, or inhalation hazard
	Inhalation hazard Respiratory or internal health
	Flammable Flammable to extremely flammable
	Oxidizing
	Explosive
	Toxic
	Compressed gas
	Keep away from sunlight Prevent exposure to sunlight and heat.

Symbol	Symbol Title and Description
	Up Store in an upright position.
	Do not freeze
	Temperature limit Upper and lower limits of temperature indicators are adjacent to the upper and lower horizontal lines.
	Handheld barcode scanner
	<i>In vitro</i> diagnostic medical device
	Contains sufficient for <n> tests Total number of IVD tests the system can perform with the IVD kit reagents appears adjacent to the symbol.
RxOnly	Prescription device (US only) Applies only to United States-registered IVD assays. CAUTION: Federal (USA) law restricts this device to sale by or on the order of a licensed healthcare professional.
	Mixing of substances Mix product before use.
	Reconstitute and mix lyophilized product before use.
	Target
	Interval
	Legal Manufacturer
	Authorized Representative in the European Community
	Use-by date Use by the designated date.
	Batch code
	Catalog number
	Recycle
	Printed with soy ink
	CE Mark


Symbol	Symbol Title and Description
	CE Mark with notified body ID number Notified body ID number can vary.
YYYY-MM-DD	Date format (year-month-day)
	Variable hexadecimal number that ensures the Master Curve and Calibrator definition values entered are valid.
	Common Units
	International System of Units
	Material
	Unique material identification number
	Name of control
	Type of control

Legal Information

Atellica is a trademark of Siemens Healthcare Diagnostics.

All other trademarks and brands are the property of their respective owners.

© 2017–2019 Siemens Healthcare Diagnostics. All rights reserved.

 Siemens Healthcare Diagnostics Inc.
511 Benedict Avenue
Tarrytown, NY 10591
USA
siemens.com/healthineers

Siemens Healthineers Headquarters

Siemens Healthcare GmbH
Henkestr. 127
91052 Erlangen
Germany
Phone: +49 9131 84-0
siemens.com/healthineers