

## HDL/LDL Cholesterol Calibrator (HDL/LDL CHOL CAL)

<b>Current Revision and Date<sup>a</sup></b> Rev. 09, 2021-11	
<b>Product Name</b>	ADVIA Chemistry HDL/LDL Cholesterol Calibrator (HDL/LDL CHOL CAL) <span style="border: 1px solid black; padding: 0 2px;">REF</span> 00309530 (B03-4763-01)
<b>Abbreviated Product Name</b>	ADVIA Chemistry HDL/LDL CHOL CAL
<b>Materials Provided</b>	3 x 0.17 g HDL/LDL CHOL CAL <span style="border: 1px solid black; padding: 0 2px;">HDL/LDL CHOL CAL</span> Calibrator lot-specific value sheet
<b>Systems</b>	ADVIA Chemistry Systems

<sup>a</sup> A vertical bar in the page margin indicates technical content that differs from the previous version.



### Intended Use

This product is intended for *in vitro* diagnostic use in the calibration of ADVIA® Chemistry systems for the Direct HDL Cholesterol and Direct LDL Cholesterol methods.

### Material Description

Material Description	Storage	Stability <sup>a</sup>
ADVIA Chemistry HDL/LDL CHOL CAL Lyophilized; reconstituted to 1.0 mL/vial Lyophilized human serum	Unopened at 2-8°C Reconstituted at 2-8°C	Until expiration date on product 3 days

<sup>a</sup> Refer to *Storage and Stability*.

For lot-specific values, refer to the lot-specific value sheet included with the calibrator kit.

### Warnings and Precautions

For *in vitro* diagnostic use.

For Professional Use.

#### CAUTION

Federal (USA) law restricts this device to sale by or on the order of a licensed healthcare professional.

Safety data sheets (SDS) available on [siemens-healthineers.com](https://www.siemens-healthineers.com).

**Warning! Potential Biohazard**

Contains human source material.

**Caution:** No known test method can ensure that products derived from human source materials will not transmit infection. These materials should be handled using good laboratory practices and universal precautions.<sup>1-3</sup>

Dispose of hazardous or biologically contaminated materials according to the practices of your institution. Discard all materials in a safe and acceptable manner and in compliance with prevailing regulatory requirements.

**Note** For information about calibrator preparation, refer to *Preparing the Calibrators*.

**Storage and Stability**

Store all calibrators in an upright position. Do not use products beyond the expiration date printed on the product labeling.

For details about product material description, storage, and stability, refer to *Material Description*.

**Performing Calibration****Calibration Frequency**

Follow government regulations or accreditation requirements for calibration frequency. Individual laboratory quality control programs and procedures may require more frequent calibration.

For information about calibration frequency, refer to the assay instructions for use.

**Preparing the Calibrators**

Prepare the calibrator using the following steps:

1. Pipette exactly 1.0 mL of distilled water into one (1) vial of the lyophilized material.
2. Allow to stand for 30 minutes.
3. Ensure homogeneity prior to use.

**Calibration Procedure**

Use the following lot-specific materials to perform calibration:

- For the assigned values and calibrator definitions, refer to the calibrator lot-specific value sheet 

CAL	LOT	VAL
-----	-----	-----

 provided. The assigned values are traceable to the standardization of the assay.
- Generate lot-specific barcode labels to use with the calibrator samples, if necessary.

For instructions about how to perform the calibration procedure, refer to the system operating instructions.

**Technical Assistance**

According to EU regulation 2017/746, any serious incident that has occurred in relation to the device shall be reported to the manufacturer and the competent authority of the EU Member State in which the user and/or patient is established.

For customer support, contact your local technical support provider or distributor.





























siemens-healthineers.com


## References

1. US Department of Health and Human Services. *Biosafety in Microbiological and Biomedical Laboratories*. 5th ed. Washington, DC: US Government Printing Office; December 2009.
2. World Health Organization. *Laboratory Biosafety Manual*. 3rd ed. Geneva: World Health Organization; 2004.
3. Clinical and Laboratory Standards Institute. *Protection of Laboratory Workers From Occupationally Acquired Infections; Approved Guideline—Fourth Edition*. Wayne, PA: Clinical and Laboratory Standards Institute; 2014. CLSI Document M29-A4.

## Definition of Symbols

The following symbols may appear on the product labeling:

Symbol	Symbol Title	Symbol	Symbol Title
	Manufacturer		Authorized representative in the European Community
	Use-by date		Batch code
	Catalog number		Contains sufficient for <n> tests
	Consult Instructions for Use		Version of Instructions for Use
	Internet URL address to access the electronic instructions for use		Revision
	<i>In vitro</i> diagnostic medical device		Unique Device Identifier
	Prescription device (US only)		CE Marking
	CE Marking with Notified Body		Keep away from sunlight
	Temperature limit		Lower limit of temperature
	Upper limit of temperature		Do not freeze
	Do not re-use		This way up
	Recycle		Caution
	Biological risks		Document face up <sup>a</sup>
	Common Units		International System of Units

Symbol	Symbol Title	Symbol	Symbol Title			
YYYY-MM-DD	Date format (year-month-day)	YYYY-MM	Date format (year-month)			
	Reconstitution volume	<table><tr><td>CAL</td><td>LOT</td><td>VAL</td></tr></table>	CAL	LOT	VAL	Calibrator lot value
CAL	LOT	VAL				
<table><tr><td>CONTROL</td><td>LOT</td><td>VAL</td></tr></table>	CONTROL	LOT	VAL	Quality control lot value		
CONTROL	LOT	VAL				

<sup>a</sup> Indicates Assay-eNote

## Legal Information

ADVIA is a trademark of Siemens Healthineers.

© 2009–2021 Siemens Healthineers. All rights reserved.



Siemens Healthcare Diagnostics Inc.  
511 Benedict Avenue  
Tarrytown, NY 10591 USA

### Siemens Healthineers Headquarters

Siemens Healthcare GmbH  
Henkestraße 127  
91052 Erlangen  
Germany  
Phone: +49 9131 84-0  
siemens-healthineers.com